

# TRACER PockeTesters

A pocket-sized ISE meter for measuring total chlorine. Use it to test pH and ORP with interchangeable flat surface sensors.

## Total Chlorine TRACER

Order Code 1740

- Read Total Chlorine from 0.00-10 ppm
- Readings are not affected by sample color or turbidity
- Automatic self calibration; extra bold display includes an analog bar graph feature; memory can store up to 15 readings
- Chlorine and pH modes also display sample temperature
- Unit identifies which probe is in use and retains calibrations
- Automatic shut-off and Low Battery indicator; uses four 3V CR-2032 batteries
- Includes 100 reagent tablets at almost half the price of similar Chlorine ISE reagents
- Follows EPA protocol for ISE methods



1740-KIT-01

## TRACER KIT with pH, Total Chlorine and ORP probes

Order Code 1740-KIT-01

Includes Tracer meter with pH, Total Chlorine and ORP probes, 100 Chlorine tablets, 50 pH 7.0 buffer tablets, tablet crusher and convenient carrying case.



**EPA  
APPROVED**  
for NPDES  
monitoring

## pH TRACER

Order Code 1741

Provided with 4, 7, and 10 pH buffer tablets.

- Rugged flat surface electrode will alert user when it's time to "RENEW"
- A "CAL" indicator shows when to recalibrate and user can select a 1, 2, or 3 point calibration
- Includes Automatic Temperature Compensation and displays temperature while showing pH result

Range:	0.00 to 14.00 pH
Temp:	32° to 149°F (0° to 65°C)
Resolution:	0.01 pH
Accuracy:	±0.01 pH

## TRACER with pH and Total Chlorine probes

Order Code 1741-KIT-01

Includes Tracer with pH and Total Chlorine probe, 100 Chlorine tablets, 50 pH tablets, tablet crusher in a convenient carrying case.



## OPTIONS

### Additional Probes

Order Code 1733 pH Sensor [0-14.00/±0.01 pH]

Order Code 1734 ORP Sensor [±999mV/±4mV]

Order Code 1732 Cl<sub>2</sub> Sensor [0-10.00/±10% of reading]

### Chlorine Test Tablets

Order Code 7044A-J

Specially formulated just for the TRACER, these deliver a precise amount of iodide for a 20 mL sample. Packages of 100.

## ORP TRACER

Order Code 1742

- High resolution to 1 mV
- Automatic self calibration

Range:	-999 to 999 mV
Resolution:	1 mV
Accuracy:	±4 mV

## TRACER KIT

with pH and ORP probes

Order Code 1742-KIT-01

Includes Tracer with pH and ORP probe, 100 Chlorine tablets, 50 pH tablets, tablet crusher in a convenient carrying case.