



Total Coliform & *E. coli* Indicator Bacteria Kit



Code 4-3616-UV

[Details on Back >](#)

Water Testing Leader Since 1919!


Total Coliform & *E. coli* Bacteria Test · 4-3616-UV

A simple 5-tube method to indicate the presence or absence of Total Coliform & *E. coli* Bacteria in drinking water. *E. coli* produces fluorescent compound.



Code 4-3616-UV

- Presumptive test for Total Coliform & *E. coli* Bacteria
- NO incubation equipment required
- Results in 44-48 hours at room temperature (70° - 85°F) or 24-hours at 110°F
- UV light source included (365 nm)
- Portable, no accessory labware required
- Ideal test for well water and coliform breakthrough in distribution systems
- Independent laboratory tested [results available upon request]



POSITIVE TEST NEGATIVE TEST

Coliform Bacteria Test · 4-3616

A simple 5-tube method to indicate the presence or absence of Total Coliform Bacteria in drinking water:

- Presumptive test for Total Coliform Bacteria
- NO incubation equipment required
- Results in 44-48 hours at room temperature (70° - 85°F)
- Portable, no accessory labware required
- Ideal test for well water and coliform breakthrough in distribution systems
- Independent laboratory tested [results available upon request]



Code 4-3616

Code	Test System	Range/Sensitivity	# of Tests [# of Reagents]	Ship Code (Wgt./lb.)
4-3616-UV	Tableted nutrient based	Presence/Absence	1 [1]	NH [1]
4-3616	Tableted nutrient based	Presence/Absence	1 [1]	NH [1]



POSITIVE TEST NEGATIVE TEST



LaMotte Company · PO Box 329 · Chestertown · Maryland · 21620 · USA
t: 800-344-3100 · 410-778-3100 · f: 410-778-6394 · www.lamotte.com